

Cleaning the Face Shield

Face shields delivered to facilities should be considered non-sterile. Consult the CDC for the latest guidance on disinfectants known to be effective against COVID-19. The following cleaning materials may be appropriate:

- Soap-water
- Isopropanol (70%) wipe

The following are not compatible with DPR 10 and **should not** be used to clean face shields:

- Hydrogen peroxide vapor sterilization
- Autoclave
- Soak in isopropanol or organic solvent